

# Bedienungsanleitung

# POSIMED®

## POSIMED®

Lagerungskissen

Lesen Sie bitte die Bedienungsanleitung vor Einsatz des Systems aufmerksam durch und beachten Sie die **Warnhinweise!**

<b>Produktname</b>	POSIMED®
<b>Lieferumfang</b>	Lagerungskissen mit PU-Bezug, Bedienungsanleitung
<b>Indikationen / Anwendungsbereiche</b>	<ul style="list-style-type: none"><li>• Dekubitusprophylaxe</li><li>• Dekubitustherapie</li><li>• Schmerztherapie</li></ul>
<b>Kontraindikationen</b>	Keine bekannt
<b>Zulässiges Patientengewicht min. - max.</b>	Für alle Patientengewichte geeignet.
<b>Material</b>	<ul style="list-style-type: none"><li>• Füllung: Polypropylen-Mikroperlen</li><li>• Bezug: PU-beschichtetes PES-Trägermaterial</li></ul>
<b>Abmessungen</b>	Siehe Label am Produkt
<b>Gewicht des Produkts</b>	0,2 - 1 kg
<b>Hilfsmittelnummer (i.S. §139 SGB V)</b>	
<b>Medizinprodukt</b>	Klasse 1, MDR 2017/745 
<b>Technische Sicherheit / Schwerentflammbarkeit</b>	Prüfung der Schwerentflammbarkeit nach DIN EN 597 I + II, BS 7175 sec 3 crib 5, BS 7177
<b>Prüfung auf toxikologische Unbedenklichkeit / Biokompatibilität</b>	DIN EN ISO 10993-5, DIN EN ISO 10993-1, DIN EN ISO 10993-10
<b>Warnhinweise</b>	Vor der ersten Anwendung muss die Vorbereitung des Systems für den ersten Einsatz durchgeführt werden (siehe "Vorbereiten des Systems für den Einsatz").
<b>Garantie</b>	5 Jahre auf die funktionellen Eigenschaften (Pro-Rata-Garantie)
<b>Wartung</b>	Keine
<b>Reinigungsanleitung (Kissenbezug)</b>	Der Bezug kann durch eine Wischdesinfektion mit handelsüblichen Desinfektionsmitteln aus Isopropanol (z.B. Bacillo) oder mit anderen geeigneten DGHM-gelisteten Produkten (z.B. ECOLAB-Reinigungstuch Incidin OxyWipe S XL) desinfiziert werden. Der Bezug kann durch Autoklavieren gemäß Richtlinien des RKIs wiederaufbereitet werden. 
<b>Reinigungsanleitung (Kern)</b>	Das gesamte Kissen kann durch Autoklavieren gemäß Richtlinien des RKIs wiederaufbereitet werden.
<b>Vorbereiten des Systems für den Einsatz</b>	<ol style="list-style-type: none"><li>1. Entnehmen Sie das Kissen aus dem Karton.</li><li>2. Das System ist jetzt einsatzbereit.</li></ol>

RKI = Robert-Koch-Institut  
DHGM = Deutsche Gesellschaft für Hygiene und Mikrobiologie  
MPG = Medizinproduktegesetz  
Autoklavieren = Fraktioniertes Vakuumverfahren bei 105 Grad C und 55 Minuten

 Hersteller im Sinne des MPG und MDR 2017/745

Funke Medical AG • Ährenfeld 10 • 46348 Raesfeld • Tel.: +49-2865-266 93 50 • Fax: +49-2865-8133  
info@funke-medical.de • www.funke-medical.de

Download: 24.06.2020 12:02

# operation manual

# POSIMED®

## POSIMED®

Positioning pillow

Please read the operating instructions carefully before using the system and heed the **warnings!**

<b>Product name</b>	<b>POSIMED®</b>
<b>scope of delivery</b>	Storage cushion with PU cover, operating instructions
<b>Indications / areas of application</b>	<ul style="list-style-type: none"><li>• Pressure ulcer prophylaxis</li><li>• Pressure ulcer therapy</li><li>• Pain therapy</li></ul>
<b>Contraindications</b>	Not known
<b>Permitted patient weight min. - Max.</b>	Suitable for all patient weights.
<b>material</b>	<ul style="list-style-type: none"><li>• Filling: polypropylene microbeads</li><li>• Cover: PU-coated PES carrier material</li></ul>
<b>Dimensions</b>	See label on the product
<b>Product weight</b>	0.2 - 1 kg
<b>Aid number (as defined in §139 SGB V)</b>	
<b>Medical device</b>	Class 1, MDR 2017/745 
<b>Technical safety / flame retardancy</b>	Flame retardancy test according to DIN EN 597 I + II, BS 7175 sec 3 crib 5, BS 7177
<b>Testing for toxicological harmlessness / biocompatibility</b>	DIN EN ISO 10993-5, DIN EN ISO 10993-1, DIN EN ISO 10993-10
<b>Warnings</b>	Before the first application, the system must be prepared for the first use (see "Preparing the system for use").
<b>guarantee</b>	5 years on the functional properties (pro-rata guarantee)
<b>maintenance</b>	No
<b>Cleaning instructions (pillowcase)</b>	The cover can be disinfected by wiping with commercially available disinfectants made from isopropanol (e.g. Bacillo) or with other suitable DGHM-listed products (e.g. ECOLAB cleaning cloth Incidin OxyWipe S XL). The cover can be reprocessed by autoclaving according to the guidelines of the RKI. 
<b>Cleaning instructions (core)</b>	The entire pillow can be reprocessed by autoclaving according to the guidelines of the RKI.
<b>Prepare the system for use</b>	<ol style="list-style-type: none"><li>1. Remove the pillow from the box.</li><li>2. The system is now ready for use.</li></ol>

RKI = Robert Koch Institute

DHGM = German Society for Hygiene and Microbiology

MPG = Medical Devices Act

Autoclaving = fractional vacuum process at 105 degrees C and 55 minutes

 **Manufacturer in the sense of the MPG and MDR 2017/745**

Funke Medical AG • Ährenfeld 10 • 46348 Raesfeld • Tel. : + 49-2865-266 93 50 • Fax: + 49-2865-8133

info@funke-medical.de • www.funke-medical.de

Download: 02.07.2020 12:30